



THC CAL 50 Urine

# Cannabinoids / *Cannabinoides* High Calibrator 50 Urine / *Calibrador Orina Alto 50* DoA Liquid Test

## Cannabinoids Calibrator High Cut-off (50 ng/mL) for Urine IVD

Store at 2 - 8°C.

### INTENDED USE

The Cannabinoids Calibrator is used for the calibration of Spinreact DoA-EIA qualitative liquid test for Cannabinoids at the high cutoff level (50ng/mL).

### REAGENTS

Ref. 933072	Cannabinoids Calibrator High Cut-off (50 ng/mL) <b>Human Urine. Sodium azide 0.5 g/l. THC</b>
-------------	--

The concentration of drug in this calibrator is:

	ng/mL
11-nor- $\Delta^9$ -THC-9-COOH	50

### PREPARATION AND STABILITY

The calibrator is ready to use. No calibrator preparation is required. The calibrator is stable until the expiration date on the label when stored tightly closed at 2-8°C and contaminations are prevented during their use.

### WARNING AND PRECAUTIONS

- This calibrator is for in vitro diagnostic use only. Harmful if swallowed.
- The calibrator contains sodium azide that may react with lead or copper plumbing to form potentially explosive metal azide. When disposing such calibrators or wastes always flush with a large volume of water to prevent azide build-up.
- Do not use the calibrator beyond their expiration dates.
- Keep all containers closed when not in use to avoid microbial contamination.
- Do not mix calibrators from different manufacturers.
- Do not freeze the calibrator.

Handle all urine specimens as if they were potentially infectious.

### PROCEDURE

The calibrator should be used in the Cannabinoids reagent for urine.

### REFERENCES

1. Urine testing for Drug of Abuse. National Institute on Drug Abuse (NIDA) Research Monograph 73, 1986.
2. European Laboratory Guidelines for Legally Defensible Workplace Drug Testing – Urine Drug Testing, EWDTS, 2002.
3. Spinreact THC EIA (Refs #930028/930030) package inserts.

### PACKAGING

Ref. 936072 **CONT.** 1x4 mL

## Calibrador *Cut-off* Alto (50 ng/mL) de Cannabinoides para Orina IVD

Conservar a 2 - 8°C.

### USO RECOMENDADO

El calibrador Cannabinoides se utiliza para la calibración del test líquido Spinreact DoA-EIA cualitativo para Cannabinoides utilizando el nivel de cutoff alto (50ng/mL).

### REACTIVOS

Ref. 933072	Calibrador <i>Cut-off</i> Alto (50 ng/mL) de Cannabinoides <b>Orina Humana. Sodio azida 0.5 g/l. THC</b>
-------------	---

La concentración de droga en este calibrator es:

	ng/mL
11-nor- $\Delta^9$ -THC-9-COOH	50

### PREPARACIÓN Y ESTABILIDAD

El calibrador está listo para su uso. No se requiere preparación del calibrador. El calibrador es estable hasta la fecha de caducidad indicada en el envase cuando se mantienen los viales bien cerrados a 2-8°C, y se evita la contaminación durante su uso.

### ADVERTENCIAS Y PRECAUCIONES

- Este calibrador es sólo para diagnóstico in vitro. Nocivo por ingestión.
- El calibrador contiene azida sódica que puede reaccionar con el cobre o plomo de las tuberías y formar componentes explosivos. En caso de eliminación de los calibradores o residuos, hacerlo con abundante agua del grifo.
- No utilizar los calibradores más allá de su fecha de caducidad.
- Mantener todos los viales cerrados cuando no están en uso para evitar contaminación microbiana.
- No mezclar calibradores de distintos fabricantes.
- No congelar los calibradores.

Manipular todas las muestras de orina como agentes potencialmente infecciosos.

### PROCEDIMIENTO

El calibrador debe ser usado en el reactivo de Cannabinoides para orina.

### REFERENCIAS

1. Urine testing for Drug of Abuse. National Institute on Drug Abuse (NIDA) Research Monograph 73, 1986.
2. European Laboratory Guidelines for Legally Defensible Workplace Drug Testing – Urine Drug Testing, EWDTS, 2002.
3. Hoja de instrucciones del kit Spinreact THC-EIA (Refs # 930028/930030).

### PRESENTACIÓN

Ref. 936072 **CONT.** 1x4 mL

